

Registrar

TENDER NOTICE

(As per G.O. No. 5400-F (Y), dated 25.06.2012)

One **laminar airflow** will be purchased by the Department of Bioinformatics, University of North Bengal. Interested suppliers may submit sealed quotations with VAT, CST, PAN details in an open envelope to the said office within 15 business days from the date of publication of the notice.

Number of laminar flow	Specification
1	Working Table S.S 304 Filter HEPA 0.3 micron of .efficiency of 99.97% down to 0.3 micron Pre filter Dry fibre washable type Standards International 209 E Blower Heavy duty Media Glass and made of pleats back in an anodized aluminium frame Velocity at the output HEPA is 90 ± 20 FPM Noise level Low Vibration Low Fluorescent Light Yes UV Light, Yes Static pressure Manometer Yes Gas cock Yes Front and side door ACRYLIC SHEET (Transparent) Front door Folding type UV Light timer 0-999min

Kindly quote with 5% GST, we will provide the Department of Scientific & Industrial Research (DSIR) Certificate.

Terms and Conditions

1. The Vendor (s) should provide the self-attested copies of VAT, CST, PAN etc failing which the quotation would not be considered.
2. Quotation should be valid for at least three months.
3. Delivery should be made within two weeks after issuing the supply orders.
4. Warranty and TC of warranty should be mentioned clearly and should not be less for a period of one year (with spare and labor) from the date of installation.
5. The University authority reserves the right to accept or reject any / all quotation (s) without assigning any reason and shall not be bound to accept the lowest quotation.

-Sd-
Registrar (Offg.)

The quotation may be addressed to

The Head
Department of Bioinformatics
University of North Bengal
Raja Rammohunpur, North Bengal
University-PO
Siliguri, District-Darjeeling, West
Bengal-734013

Advt.No:- 98/R-2022

Date 24-03-2022.