

Ref. No.: ZOO/0025/2022

Date: 20.01.2022

NOTICE

Sealed quotations are invited for the purchase of UV Transilluminator as per the specifications provided. Quotations must accompany GST and PAN photocopies, duly stamped and signed along with valid certificates. Quotations should remain valid for 90 days from the last date of submission. All documents must reach in sealed cover to Dr. Dhiraj Saha, Department of Zoology, University of North Bengal on or before 31<sup>st</sup> January, 2022 at 5 pm.

Sd/-

Dr. Dhiraj Saha

Principal Investigator

DST SERB Project

Project No. EEQ/2019/000197

# University of North Bengal

Accredited by NAAC with Grade A

**DEPARTMENT OF ZOOLOGY**

DST-FIST & UGC-SAP Sponsored



समानो मन्त्र समिति समानो

P.O. North Bengal University  
Raja Rammohunpur, Dt. Darjeeling  
West Bengal, India, PIN-734013

## Specification of UV Transilluminator

- Gel view area: 20×20 cm
- Dual Wavelength-312nm and 365nm
- Side-by-side UV/white filter sections
- Pure -UV quartz filter to view DNA and RNA gels,
- High intensity UV tubes, ultra-safe UV blocking lid
- Limit of detection: 0.1 to 10 ng of DNA
- Camera holder: Size: 20×20 cm

*D. Saha*  
20.01.22

Dr. Dhiraj Saha, Ph.D.  
Principal Investigator  
DST SERB Project  
No. EEQ/2019/000197